



CERTIFICATE OF ANALYSIS

RESTORE

1700-0140

Product: RESTORE

Part Number: 1700-0140

Lot Number: 405020

Production Date: 20 May 2014

Expiration Date: 20 May 2016

Primary Application: Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

Acceptance Criteria:

1. pH = 1.30 ± 0.05 @ 25 ± 0.4 °C (Y/N)

y

Analysis Certified by: Jim Murphy

Date: 10 Jun 2014

President

Title