



CERTIFICATE OF ANALYSIS

Fusayama / Meyer Artificial Saliva 1700-0301

Product: Artificial Saliva for testing of biodegradability of dental metal alloys

Part Number: 1700-0301

Lot Number: 011041

Production Date: November 18, 2020

Expiration Date: November 18, 2021

Primary Application: Product Testing

Acceptance Criteria:

1. pH = 4.9 ± 0.1 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Keep refrigerated

Analysis Certified by: Jim Murphy
President

Title

Date: December 3, 2020

Rev. A, 03/26/14