



## CERTIFICATE OF ANALYSIS

**Fusayama / Meyer Artificial Saliva  
1700-0301**

**Product:** Artificial Saliva for testing of biodegradability of dental metal alloys

**Part Number:** 1700-0301

**Lot Number:** 712011

**Production Date:** 06 Dec 2017

**Expiration Date:** 06 Dec 2018

**Primary Application:** Product Testing

**Acceptance Criteria:**

1. pH =  $4.9 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)

y

**Storage Conditions:** Keep refrigerated

**Analysis Certified by:** Jim Murphy  
President

Title

**Date:** December 8, 2017

Rev. A, 03/26/14