



Certificate of Analysis

Product Artificial Saliva for AFNOR NF S91-141
Part Number 1700-0302
Primary Application Artificial Saliva for testing of biodegradability of dental metal alloys
Volume 200 mL
Lot Number 2203015
Production Date 08 March 2022
Expiration Date 08 March 2023
Storage Conditions Refrigerated at 4 °C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 6.9 ± 0.1 @ 25 ± 0.4 °C :.....PASS
3. Other components gravimetrically-certified:.....PASS

Components

Potassium Thiocyanate Potassium Chloride
Potassium Phosphate Sodium Bicarbonate
Sodium Chloride

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: Maria Ofitserova, Ph.D.

Date: 18 March 2022

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED