



## Certificate of Analysis

**Product** Artificial Saliva for AFNOR NF S91-141  
**Part Number** 1700-0302  
**Primary Application** Artificial Saliva for testing of biodegradability of dental metal alloys  
**Volume** 200 mL  
**Lot Number** 2310063  
**Production Date** 30 Oct 2023  
**Expiration Date** 30 Oct 2024  
**Storage Conditions** Refrigerated at 4 °C  
**Acceptance Criteria**

1. Appearance:.....PASS  
2. pH =  $6.9 \pm 0.1$  @  $25 \pm 0.4$  °C :.....PASS  
3. Other components gravimetrically-certified:.....PASS

### Components

Potassium Thiocyanate Potassium Chloride  
Potassium Phosphate Sodium Bicarbonate  
Sodium Chloride

\*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

**Analysis Certified By:** David Mazawa

**Date:** 31 Oct 2023

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED