



## CERTIFICATE OF ANALYSIS

### Artificial Saliva for ARNOF NF S91-141 1700-0302

**Product:** Artificial Saliva for testing of biodegradability of dental metal alloys according to ARNOF NF S91-141

**Part Number:** 1700-0302

**Lot Number:** 603013

**Production Date:** 07 Mar 2016

**Expiration Date:** 07 Mar 2017

**Primary Application:** Product Testing

#### Acceptance Criteria:

pH =  $6.9 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)  
measured before adding Sodium Bicarbonate

y

**Storage Conditions:** Keep refrigerated

**Analysis Certified by:** Jim Murphy  
President  
Title

**Date:** March 8, 2016

Rev. A, 03/26/14