



CERTIFICATE OF ANALYSIS

Artificial Saliva for DIN 53160 1700-0303

Product: Artificial Saliva for testing of colorfastness according to method DIN53160

Part Number: 1700-0303

Lot Number: 409005

Production Date: 04 Aug 2014

Expiration Date: 04 Aug 2015

Primary Application: Product Testing

Acceptance Criteria (after combining Part 1 and Part 2):

1. pH = 6.8 ± 0.1 @ 25 ± 0.4 °C (Y/N) y

Storage Conditions: Keep refrigerated

Analysis Certified by: Jim Murphy **Date:** 15 Sep 2014
President
Title

Rev. A, 03/26/14