



## Certificate of Analysis

**Product** Artificial Saliva for Medical and Dental Research  
**Part Number** 1700-0305  
**Primary Application** Research and product testing  
**Volume** 200 mL  
**Lot Number** 2307040  
**Production Date** 25 July 2023  
**Expiration Date** 25 July 2024  
**Storage Conditions** Dry at 8° to 25°C

### Acceptance Criteria

1. Appearance:.....PASS  
2. pH =  $6.75 \pm 0.05$  @  $25 \pm 0.4^\circ\text{C}$  :.....PASS  
3. Other components gravimetrically-certified:.....PASS

### Components

DI Water	Potassium Phosphate	Preservative
Na Carboxymethyl Cellulose	Magnesium Chloride	
Potassium Chloride	Calcium Chloride	

\*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

**Analysis Certified By:** David Mazawa

**Date:** 27 July 2023

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED