



CERTIFICATE OF ANALYSIS

Fusayama / Meyer Artificial Saliva – Stabilized 1700-0309

Product: Fusayama / Meyer Artificial Saliva – Stabilized

Part Number: 1700-0309

Lot Number: 003019

Production Date: March 10, 2020

Expiration Date: March 10, 2021

Primary Application: Product Testing

Acceptance Criteria:

1. pH = 4.9 ± 0.1 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: 8 °C – 25 °C

Analysis Certified by: Rebecca Smith
Vice President

Title

Date: 3/18/2020

Rev. A, 02/14/2018