



Certificate of Analysis

Product Artificial Saliva ASTM E2720-16/ASTM E2721-16 with Mucin, 2-part
Part Number 1700-0317
Primary Application Product testing and research
Volume 200 mL
Lot Number 2112002
Production Date 02 Dec 2021
Expiration Date 02 Dec 2022
Storage Conditions Refrigerated at 4 °C

Acceptance Criteria

1. Appearance:.....PASS
 2. pH = 7.0 ± 0.2 @ 25 ± 0.4°C :.....PASS
 3. Other components gravimetrically-certified:.....PASS

Components

DI Water	Potassium Chloride	Magnesium Chloride	Ammonium Chloride
Sodium Chloride	Potassium Thiocyanate	Sodium Bicarbonate	Mucin
Potassium Phosphate	Calcium Chloride	Urea	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 29 Dec 2021

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED