



## Certificate of Analysis

**Product** Artificial Saliva for DIN 53160, GB/T 18886, and SN/T 1058  
**Part Number** 1700-0323  
**Primary Application** Testing of colorfastness  
**Volume** 950 mL  
**Lot Number** 2210015  
**Production Date** 12 Oct 2022  
**Expiration Date** 12 Nov 2022  
**Storage Conditions** Refrigerated at 4 °C  
**Acceptance Criteria**

1. Appearance:.....PASS  
2. pH =  $6.8 \pm 0.1$  @  $25 \pm 0.4$  °C :.....PASS  
3. Other components gravimetrically-certified:.....PASS

### Components

|                     |                     |
|---------------------|---------------------|
| Magnesium Chloride  | Sodium Chloride     |
| Calcium Chloride    | Potassium Chloride  |
| Potassium Phosphate | Potassium Carbonate |

\*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

**Analysis Certified By:** David Mazawa

**Date:** 12 Oct 2022

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED