



Certificate of Analysis

Product Artificial Perspiration according to ISO 105-B07 / ISO 105-E04, pH 5.5, Stabilized
Part Number 1700-0507
Primary Application Determination of color fastness to perspiration (fabric).
Volume 200 mL
Lot Number 2510059
Production Date 30 Oct 2025
Expiration Date 30 Oct 2026
Storage Conditions Dry at 8° to 25°C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 5.5 ± 0.1 @ $25 \pm 0.4^{\circ}\text{C}$:.....PASS
3. Other components gravimetrically-certified:.....PASS

Components


Water	Sodium dihydrogen -
Sodium Chloride	orthophosphate dihydrate
L-Histidine	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 03 Nov 2025

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED