



Certificate of Analysis

Product Artificial Perspiration, ISO 3160, Custom pH, Stabilized
Part Number 1700-0526
Primary Application Determination of corrosion resistance
Volume 200 mL
Lot Number 2312016
Production Date 07 Dec 2023
Expiration Date 07 Dec 2024
Storage Conditions Dry at 8° to 25°C
Acceptance Criteria

1. Appearance:.....PASS
 2. pH = 8.6 ± 0.1 @ 25 ± 0.4°C :.....PASS
 3. Other components gravimetrically-certified:.....PASS

Components

DI Water	Acetic Acid	Preservative
Sodium Chloride	Ammonium Chloride	
DL-Lactic Acid	Urea	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 12 Dec 2023

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED