



CERTIFICATE OF ANALYSIS

Artificial Eccrine Perspiration – Stabilized 1700-0531

Product: Artificial Eccrine Perspiration – Stabilized, 950mL

Part Number: 1700-0531

Lot Number: 808050

Production Date: 28 Aug 2018

Expiration Date: 28 Aug 2020

Primary Application: Artificial Eccrine Perspiration is a solution of amino acids, salts and metabolites that are found in human eccrine perspiration. It is used in sweat testing of products.

Acceptance Criteria:

1. pH = 4.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)

y

2. Concentrations of amino acids and other components meet specifications (Y/N):

y

Analysis Certified by: Jim Murphy

Date: September 6, 2018

President

Title