



Certificate of Analysis

Product Artificial Perspiration according to AATCC Test Method 15 (Stabilized)
Part Number 1700-0555
Primary Application Determination of colorfastness to perspiration.
Volume 19.8 L
Lot Number 2110041
Production Date 18 October 2021
Expiration Date 18 October 2022
Storage Conditions Dry at 8° to 25° C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 4.3 ± 0.1 @ $25 \pm 0.4^{\circ}\text{C}$:.....PASS
3. Other components gravimetrically-certified:.....PASS

Components

DI Water	L-Histidine
Sodium Chloride	DL-Lactic Acid
Sodium Phosphate dibasic	Preservative

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: Maria Ofitserova, Ph.D

Date: 26 October 2021

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED