PICKERING LABORATORIES

Certificate of Analysis

Product	Artificial Perspiration according to AATCC Test Method 15 (Stabilized)
Part Number	1700-0555
Primary Application	Determination of colorfastness to perspiration.
Volume	19.8 L
Lot Number	2407027
Production Date	16 July 2024
Expiration Date	16 July 2025
Storage Conditions	Dry at 8° to 25° C

Acceptance Criteria

1. Appearance: ..... PASS
2. pH = 4.3 ± 0.1 @ 25 ± 0.4°C : ..... PASS
3. Other components gravimetrically-certified: ..... PASS

Components

DI Water	L-Histidine
Sodium Chloride	DL-Lactic Acid
Sodium Phosphate dibasic	Preservative

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa **Date:** 30 July 2024

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.

GUARANTEED CHEMISTRY CERTIFIED

1280 Space Park Way | Mountain View CA 94043 | USA | 800-654-3330 | 650-694-6700
www.pickeringtestsolutions.com | support@pickeringlabs.com