



## CERTIFICATE OF ANALYSIS

### Artificial Urine DIN EN 1616:1999, Stabilized 1700-0558

**Product:** Artificial Urine DIN EN 1616:1999, stabilized

**Part Number:** 1700-0558

**Lot Number:** 2101043

**Production Date:** January 28, 2021

**Expiration Date:** January 28, 2022

**Primary Application:** For testing of urethral catheters.

**Acceptance Criteria:**

1. pH =  $6.6 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)

Y

**Storage Conditions:** Dry at 8 ° to 25° C

**Analysis Certified by:** Jim Murphy  
President

Title

**Date:** February 26, 2021

Rev. A, 05/14/19