



## CERTIFICATE OF ANALYSIS

### Artificial Urine DIN EN 1616:1999, Stabilized 1700-0558

**Product:** Artificial Urine DIN EN 1616:1999, stabilized

**Part Number:** 1700-0558

**Lot Number:** 2309058

**Production Date:** 28 Sep 2023

**Expiration Date:** 28 Sep 2024

**Primary Application:** For testing of urethral catheters.

**Acceptance Criteria:**

1. pH =  $6.6 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)

Y

**Storage Conditions:** Dry at 8 ° to 25° C

**Analysis Certified by:** David Mazawa

**Date:** 03 Oct 2023

**Title:** Technical Support Chemist