



## Certificate of Analysis

**Product** Artificial Urine DIN EN 1616:1999, Stabilized  
**Part Number** 1700-0558  
**Primary Application** Testing of urethral catheters  
**Volume** 200 mL  
**Lot Number** 2502044  
**Production Date** 26 February, 2025  
**Expiration Date** 26 February, 2026  
**Storage Conditions** Dry at 8° to 25°C  
**Acceptance Criteria**

1. Appearance:.....PASS  
2. pH =  $6.6 \pm 0.1$  @  $25 \pm 0.4^{\circ}\text{C}$  :.....PASS  
3. Other components gravimetrically-certified:.....PASS

### Components

Water	Creatinine Hydrochloride
Urea	Ammonium Chloride
Sodium Chloride	Potassium Phosphate
Sodium Sulphite	Preservative
Sodium Phosphate	

\*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

**Analysis Certified By:** Maria Ofitserova, Ph.D.

**Date:** 27 February, 2025

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED