



## Certificate of Analysis

**Product** Perspiration according to DIN 53160, BS EN 1811:2011, Buffered, Stabilized  
**Part Number** 1700-0568  
**Primary Application** Colorfastness of articles for common use.  
**Volume** 200 mL  
**Lot Number** 2308058  
**Production Date** 24 Aug 2023  
**Expiration Date** 24 Aug 2024  
**Storage Conditions** Dry at 8° to 25°C  
**Acceptance Criteria**

1. Appearance:.....PASS  
2. pH =  $6.50 \pm 0.05$  @  $25 \pm 0.4^{\circ}\text{C}$  :.....PASS  
3. Other components gravimetrically-certified:.....PASS

### Components

Water	Lactic Acid
Sodium Chloride	Preservative
Urea	Sodium Phosphate Dibasic

\*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

**Analysis Certified By:** David Mazawa

**Date:** 28 Aug 2023

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED