



Certificate of Analysis

Product Sweat DIN 53160-2:200-10, BS EN 1811:2011 Buffered, Not Stabilized, Custom pH
Part Number 1700-0569
Primary Application Determination for release of nickel
Volume 200 mL
Lot Number 2410040
Production Date 21 Oct 2024
Expiration Date 21 Oct 2025
Storage Conditions Frozen at - 20°C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 6.5 ± 0.1 @ $25 \pm 0.4^\circ\text{C}$:.....PASS
3. Other components gravimetrically-certified:.....PASS

Components

| | |
|-----------------|--------------------------|
| DI Water | Sodium Phosphate Dibasic |
| Sodium Chloride | Urea |
| DL-Lactic Acid | |

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 22 Oct 2024

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED