



Certificate of Analysis

Product Sweat DIN 53160-2:200-10, BS EN 1811:2011 Buffered, Not Stabilized, Custom pH
Part Number 1700-0569
Primary Application Determination for release of nickel
Volume 200 mL
Lot Number 2410041
Production Date 21 Oct 2024
Expiration Date 21 Oct 2025
Storage Conditions Frozen at - 20°C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 8.0 ±0.1 @ 25±0.4°C :.....PASS
3. Other components gravimetrically-certified:.....PASS

Components

DI Water	Sodium Phosphate Dibasic
Sodium Chloride	Urea
DL-Lactic Acid	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 22 Oct 2024

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED