



Certificate of Analysis

Product Artificial Perspiration DIN 53160, BS EN 1811, Buffered, Custom pH - Stabilized
Part Number 1700-0570
Primary Application Product testing
Volume 200 mL
Lot Number 2601034
Production Date 20 Jan 2026
Expiration Date 20 Jan 2027
Storage Conditions Dry at 8° to 25°C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 6.5 ± 0.1 @ $25 \pm 0.4^\circ\text{C}$:.....PASS
3. Other components gravimetrically-certified:.....PASS

Components

DI Water	Lactic Acid
Sodium Chloride	Preservative
Urea	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 20 Jan 2026

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED