



## Certificate of Analysis

**Product** Artificial Urine Stabilized  
**Part Number** 1700-0602  
**Primary Application** Product testing  
**Volume** 950 mL  
**Lot Number** 2302025  
**Production Date** 15 February 2023  
**Expiration Date** 15 February 2024  
**Storage Conditions** Dry at 8° to 25°C

### Acceptance Criteria

1. Appearance:.....PASS  
2. pH =  $6.5 \pm 0.1$  @  $25 \pm 0.4^\circ\text{C}$  :.....PASS  
3. Other components gravimetrically-certified:.....PASS

### Components

D.I. Water	Creatinine	Citric Acid	Magnesium Sulfate
Sodium Chloride	Calcium Chloride	Peptone	Preservative
Sodium Bicarbonate	Potassium Phosphate	Yeast Extract	
Urea	Ammonium Chloride	Iron Sulfate	
Uric Acid	Lactic Acid	Sodium Sulfate	

\*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

**Analysis Certified By:** Maria Ofitserova, Ph.D.

**Date:** 17 February 2023

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED