



## CERTIFICATE OF ANALYSIS

### Artificial Urine DIN EN 1616:1999 1700-0603

**Product:** Artificial Urine prepared according to DIN EN 1616:1999 standard procedure

**Part Number:** 1700-0603

**Lot Number:** 2509005

**Production Date:** 03 Sep 2025

**Expiration Date:** 03 Sep 2026

**Primary Application:** For testing of urethral catheters.

#### Acceptance Criteria:

1. pH =  $6.6 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)

Y

**Storage Conditions:** Keep frozen

**Analysis Certified by:** David Mazawa

**Title:** QC Manager

**Date:** 10 Sep 2025