



CERTIFICATE OF ANALYSIS

Synthetic Blood ISO 16603-2004

1700-0839

Product: Synthetic Blood ISO 16603-2004

Part Number: 1700-0839

Lot Number: 2409027

Production Date: 18 Sep 2024

Expiration Date: 18 Sep 2025

Primary Application: Product Testing

Acceptance Criteria:

1. Surface tension (Du Noüy method) = 0.042 ± 0.002 N/m @ $20 \pm 1^\circ\text{C}$ (Y/N) Y

2. Dynamic Viscosity = 2.7 ± 0.3 mPa*s @ $20 \pm 1^\circ\text{C}$ (Y/N) Y

3. pH = 7.3 ± 0.1 @ $25 \pm 0.4^\circ\text{C}$ (Y/N) Y

Storage Conditions: Dry at 8 to 25°C

Analysis Certified by: David Mazawa **Date:** 24 Sep 2024
Title: QC Manager