



CERTIFICATE OF ANALYSIS

KO2 ALKION Eluant

1700-1102

Product: KO2 ALKION Eluant

Part Number: 1700-1102

Lot Number: 009050

Production Date: September 23, 2020

Expiration Date: September 23, 2022

Primary Application:

1. Eluant for ALKION cation-exchange column.

Acceptance Criteria:

1. pH = 13.0 ± 0.1 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy
President
Title

Date: October 13, 2020