



CERTIFICATE OF ANALYSIS

KO2 ALKION Eluant

1700-1102

Product: KO2 ALKION Eluant
Part Number: 1700-1102
Lot Number: 103005
Test Date: 01 MAR 2011
Expiration Date: 01 MAR 2013

Primary Application:

1. Used in aminoglycoside antibiotics analyses.

Acceptance Criteria:

Actual Values:

1. pH = 13 ± 0.10 @ 25 ± 0.4 °C
2. Conductivity = 25 ± 0.2 mS @ 25 ± 0.4 °C

13.020
25.06

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Young
VP OPR

Title

Date: 02 MAR 2011

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE