



CERTIFICATE OF ANALYSIS

KO2 ALKION Eluant

1700-1102

Product: KO2 ALKION Eluant

Part Number: 1700-1102

Lot Number: 202038

Production Date: 27 Feb 2012

Expiration Date: 27 Feb 2014

Primary Application:

1. Eluant for ALKION cation-exchange column.


Acceptance Criteria:

1. pH = 13.0 ± 0.1 @ 25 ± 0.4 °C (Y/N)

4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:


Title

Date:

05 MAR 2012

ACCEPTED

PICKERING LABORATORIES, INC
QUALITY ASSURANCE