



CERTIFICATE OF ANALYSIS

Sodium Phosphate Eluant for Mycotoxins Analysis

1700-1108

Product: Eluant for Aflatoxins analysis

Part Number: 1700-1108

Lot Number: 401027

Production Date: 28 Jan 2014

Expiration Date: 28 Jan 2015

Primary Application: Aflatoxins analysis

Acceptance Criteria:

1. pH = 3.52 ± 0.05 @ 25 ± 0.4 °C (Y/N)

y

2. Well separated Aflatoxins peaks (Y/N)

y

Sample: Aflatoxins test mixture

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: Jim Murphy
President

Title

Date: 04 Feb 2014