



## CERTIFICATE OF ANALYSIS

### Sodium Phosphate Eluant for Mycotoxins Analysis

1700-1108

**Product:** Eluant for Aflatoxins analysis

**Part Number:** 1700-1108

**Lot Number:** 511025

**Production Date:** 12 Nov 2015

**Expiration Date:** 12 Nov 2016

**Primary Application:** Aflatoxins analysis

#### Acceptance Criteria:

1. pH =  $3.52 \pm 0.05$  @  $25 \pm 0.4$  °C (Y/N)

y

2. Well separated Aflatoxins peaks (Y/N)

y

**Sample:** Aflatoxins test mixture

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:** Jim Murphy  
President

Title

**Date:** November 16,2015