



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis

1700-1121 (239443)

Product: Beckman, Lithium Eluant for Amino Acid Analysis

Part Number: 1700-1121

Lot Number: 104019

Production Date: 29 APR 2011

Expiration Date: 27 APR 2012

Primary Application: Lithium based amino acid analysis in the Beckman System 6300 Amino Acids Analyzer.

Acceptance Criteria:

1. pH = 2.75 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Conductivity: 9.65 ± 0.05 mS @ 25 ± 0.4 °C (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Mumb
VP Ops
Title

Date: 03 MAY 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE