



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acid Analysis

1700-1122 (239444) Li2

Product: Beckman, Lithium Eluant for Amino Acid Analysis

Part Number: 1700-1122

Lot Number: 103013

Production Date: 09 MAR 2011

Expiration Date: 10 MAR 2012

Primary Application: Lithium based amino acid analysis in the Beckman System 6300 Amino Acids Analyzer.

Acceptance Criteria:

1. pH = 3.60 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Conductivity: 17.250 ± 0.05 mS @ 25 ± 0.4 °C (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Lynch
VP OPR

Title

Date: 12 Apr 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE