



# CERTIFICATE OF ANALYSIS

## Lithium Eluant for Amino Acids Analysis

1700-1125

Product: Lithium Eluant for Amino Acid Analysis

Part Number: 1700-1125

Lot Number: 201015

Production Date: 12 JAN 2012

Expiration Date: 10 JAN 2014

Primary Application: Lithium-based amino acids analysis.


### Acceptance Criteria:

1. pH =  $2.80 \pm 0.01$  @  $25 \pm 0.4$  °C (Y/N)
2. Resolution meets specifications:
  - Threonine/Serine (Y/N)
  - Asparagine/ Glutamic Acid (Y/N)
  - Glycine, Alanine, Citrulline (Y/N)
3. Elution time of Aspartic Acid meets specification (Y/N)
4. Baseline: no peaks (Y/N)

Y  
Y  
Y  
Y  
Y  
Y

Calibration Standard: Lithium Calibration Standard, PN: 1700-0170

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:   
Title

Date: 28 FEB 2012

# ACCEPTED

PICKERING LABORATORIES, INC.

QUALITY ASSURANCE

Rev. 12, 10/07/2011

P:\Documentation\Chemical Documentation\1700-1125