



CERTIFICATE OF ANALYSIS

Lithium Eluant for Amino Acids Analysis

1700-1125

Product: Lithium Eluant for Amino Acid Analysis

Part Number: 1700-1125

Lot Number: 2201108

Production Date: February 28, 2022

Expiration Date: February 28, 2024

Primary Application: Lithium-based amino acids analysis.

Acceptance Criteria:

1. pH = 2.80 ± 0.01 @ 25 ± 0.4 °C (Y/N)	<u>Y</u>
2. Resolution meets specifications:	
• Threonine/Serine (Y/N)	<u>Y</u>
• Asparagine/ Glutamic Acid (Y/N)	<u>Y</u>
• Glycine, Alanine, Citrulline (Y/N)	<u>Y</u>
3. Elution time of Aspartic Acid meets specification (Y/N)	<u>Y</u>
4. Baseline: no peaks (Y/N)	<u>Y</u>

Calibration Standard: Lithium Calibration Standard, **PN: 1700-0170**

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** March 4, 2022
President
Title