



Certificate of Analysis

Product

Part Number

Primary Application Product Testing

Volume

Lot Number

Production Date

Expiration Date

Storage Conditions

Acceptance Criteria

1. Appearance:.....PASS
2. Kinematic Viscosity 3.8+/- 0.3 cSt @ 20+/- 1 °C:.....PASS
3. Surface Tension measured by capillarity method 61+/- 1 dyn/cm @ 20+/-1 °C*:.....PASS
* equivalent to surface tension of 0.042 N/m measure by DuNouy ring method as per ASTM F1862/F1862M-17
4. Other components gravimetrically-certified:.....PASS

Components

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By:

Date:

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED