



CERTIFICATE OF ANALYSIS

**Thiofluor
3700-2000**

Product: Thiofluor
Part Number: 3700-2000
Lot Number: 103001
Production Date: 03 MAR 2011

Primary Application: Reagent for Post column analysis

Acceptance Criteria:

1. Color : White (Y/N)
2. Texture: Dry Crystals (Y/N)
3. Colorless when dissolved in CB910 (Y/N)
4. Carbamates peak areas meet specification (Y/N)

Y
Y
Y
Y

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: *Jim Murphy* **Date:** 14 MAR 2011

VP Ops
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE