



CERTIFICATE OF ANALYSIS

Thiofluor
3700-2000

Product: Thiofluor
Part Number: 3700-2000
Lot Number:
Production Date:
Expiration Date:

203026
27 MAR 2012
21 MAR 2015

Primary Application: Reagent for Post column analysis

Acceptance Criteria:

1. Color : White (Y/N)
2. Texture: Dry Crystals (Y/N)
3. Colorless when dissolved in CB910 (Y/N)
4. Carbamates peak areas meet specification (Y/N)
5. Baseline noise (0 +/- 25 mvolts) /(Y/N)

Y
Y
Y
Y
Y

Sample: Carbamate Test Mixture, PN: 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by:

Jim Payne
VP OpR
Title

Date:

30 MAR 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE