



## CERTIFICATE OF ANALYSIS

**Thiofluor**  
**3700-2000**

**Product:** Thiofluor  
**Part Number:** 3700-2000  
**Lot Number:** 204020  
**Production Date:** 23 Apr 2012  
**Expiration Date:** 23 Apr 2015

**Primary Application:** Reagent for Post column analysis

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. Color : White (Y/N)                            | <u>y</u> |
| 2. Texture: Dry Crystals (Y/N)                    | <u>y</u> |
| 3. Colorless when dissolved in CB910 (Y/N)        | <u>y</u> |
| 4. Carbamates peak areas meet specification (Y/N) | <u>y</u> |
| 5. Baseline noise (0 +/- 25 mvolts) /(Y/N)        | <u>y</u> |

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** Jim Murphy **Date:** 10 May 2012  
VP Operations  
**Title**