CERTIFICATE OF ANALYSIS

Thiofluor
3700-2000

Product: Thiofluor
Part Number: 3700-2000
Lot Number: 209033
Production Date: 15 Sep 2012
Expiration Date: 15 Sep 2015

Primary Application: Reagent for Post column analysis

Acceptance Criteria:

1.	Color : White (Y/N))	**y**
2.	Texture: Dry Crystals (Y/N)	**y**
3.	Colorless when dissolved in CB910 (Y/N)	**y**
4.	Carbamates peak areas meet specification (Y/N)	**y**
5.	Baseline noise (0 +/- 25 mvolts) /(Y/N)	**y**

Sample: Carbamate Test Mixture, **PN**: 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: Jim Murphy **Date:** 10 OCt 2012
VP Operations
Title

1280 Space Park Way · Mountain View CA 94043 · USA
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · www.pickeringlabs.com · support@pickeringlabs.com