



CERTIFICATE OF ANALYSIS

Thiofluor
3700-2000

Product: Thiofluor
Part Number: 3700-2000

Lot Number: 2507049
Production Date: 30 July 2025
Expiration Date: 30 July 2028

Primary Application: Reagent for Post column analysis

Acceptance Criteria:

1. Color : White (Y/N))	<u>Y</u>
2. Texture: Dry Crystals (Y/N)	<u>Y</u>
3. Colorless when dissolved in CB910 (Y/N)	<u>Y</u>
4. Carbamates peak areas meet specification (Y/N)	<u>Y</u>
5. Baseline noise (0 +/- 25 mvolts) /(Y/N)	<u>Y</u>

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: David Mazawa **Date:** 14 Aug 2025
Title: QC Manager