



## CERTIFICATE OF ANALYSIS

### Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis using EPA Method 531.2

**CB130.2**

**Product:** Hydrolysis Reagent for Post-column analysis of Carbamate Pesticides

**Part Number:** CB130.2

**Lot Number:** 2602037

**Production Date:** 19 Feb 2026

**Expiration Date:** 19 Feb 2028

**Primary Application:** Post-column Carbamate Pesticides analysis.

**Acceptance Criteria:**

- |                                                   |          |
|---------------------------------------------------|----------|
| 1. pH = $12.6 \pm 0.1$ @ $25 \pm 0.4$ °C (Y/N)    | <u>Y</u> |
| 2. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)      | <u>Y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>Y</u> |

**Sample:** Carbamate Test Mixture, **PN** 1700-0063,

**Storage Conditions:** Dry at 8 to 25 °C

|                               |                     |              |                    |
|-------------------------------|---------------------|--------------|--------------------|
| <b>Analysis Certified by:</b> | <u>David Mazawa</u> | <b>Date:</b> | <u>25 Feb 2026</u> |
| <b>Title:</b>                 | <u>QC Manager</u>   |              |                    |