



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis reagent for post-column analysis of Carbamate pesticides

Part Number: CB130

Lot Number: 410032

Production Date: 29 Oct 2014

Expiration Date: 29 Oct 2016

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

- | | |
|---|----------|
| 1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) | <u>y</u> |
| 2. Baseline noise (0 ± 25 mvolts) (Y/N) | <u>y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>y</u> |

Sample: Carbamate Test Mixture, **PN:** 1700-0063.

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: Jim Murphy **Date:** 29 Oct 2014
President
Title