



## CERTIFICATE OF ANALYSIS

### Carbamate Analysis Hydrolysis Reagent

CB130

**Product:** Carbamate Analysis Hydrolysis Reagent

**Part Number:** CB130

**Lot Number:** 806001

**Test Start Date:** 02 JUN 2010

**Expiration Date:** 01 JUN 2012

**Primary Application:** Carbamate analysis.

**Acceptance Criteria:**

- |   |               |
|---|---------------|
| 1. pH = $12.5 \pm 0.1$ @ $25 \pm 0.4$ °C          | <u>12.490</u> |
| 2. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)      | <u>y</u>      |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>y</u>      |

**Calibration Standard:** Carbamate Test Standard, PN 1700-0063, LN 908045

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:**  **Date:** 04 JUN 2010

Title

Rev. 7, 9/16/08

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE