



CERTIFICATE OF ANALYSIS

Carbamate Analysis Hydrolysis Reagent

CB130

Product: Carbamate Analysis Hydrolysis Reagent

Part Number: CB130

Lot Number: 009027

Test Start Date: 14 Sep 2010

Expiration Date: 13 Sep 2012


Primary Application: Carbamate analysis.

Acceptance Criteria:

1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C 12.550
2. Baseline noise (0 ± 25 mvolts) (Y/N) Y
3. Twelve appropriate peaks in chromatogram (Y/N) Y

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  Date: 17 Sep 2010
VP OPR
Title

Rev. 7, 9/16/08

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE