



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis Reagent for post-column analysis of Carbamate pesticides

Part Number: CB130

Lot Number: 102012

Expiration Date: 11 FEB 2013

Test Start Date: 14 FEB 2011

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria:

1. pH = $12.5 \pm 0.1 @ 25 \pm 0.4 ^\circ\text{C}$ = 12.560
2. Baseline noise (0 ± 25 mvolts) (Y/N) = Y
3. Twelve appropriate peaks in chromatogram (Y/N) = Y

Sample: Carbamate Test Mixture, **PN:** 1700-0063, **LN:** 908045

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by:

Title: V P OPR

Date: 17 FEB 2011

Rev. 7, 9/16/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE