



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis Reagent for post-column analysis of Carbamate pesticides
Part Number: CB130
Lot Number: 102028
Expiration Date: 18 Feb 2013
Test Start Date: 22 Feb 2011


Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria:

1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C 12.56
2. Baseline noise (0 ± 25 mvolts) (Y/N) Y
3. Twelve appropriate peaks in chromatogram (Y/N) Y

Sample: Carbamate Test Mixture, **PN:** 1700-0063, **LN:** 908045

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by:  **Date:** 02 MAR 2011
VP Oper
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE