



# CERTIFICATE OF ANALYSIS

## Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

**CB130**

**Product:** Hydrolysis reagent for post-column analysis of Carbamate pesticides

**Part Number:** CB130

**Lot Number:** 106 030

**Production Date:** 24 Jun 2011

**Expiration Date:** 24 Jun 2013

**Primary Application:** Post-column analysis of Carbamate pesticides.

### Acceptance Criteria

1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N) 4
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) 4
3. Twelve appropriate peaks in chromatogram (Y/N) 4

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063.

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:** *Jim Murphy*

**Date:** 28 Jun 2011

*VP Ops*  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC  
QUALITY ASSURANCE