



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis reagent for post-column analysis of Carbamate pesticides

Part Number: CB130

Lot Number: 107030

Production Date: 27 July 2011

Expiration Date: 27 Jul 2013

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Baseline noise (0 ± 25 mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

Y
Y
Y

Sample: Carbamate Test Mixture, **PN:** 1700-0063.

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: _____

Jim Murphy
VP OPR
Title

Date: _____

20 Sep 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. A, 04/08/2011

P:\Documentation\Chemical Documentation\CB130