



## CERTIFICATE OF ANALYSIS

### Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

**CB130**

**Product:** Hydrolysis reagent for post-column analysis of Carbamate pesticides

**Part Number:** CB130

**Lot Number:** 109008

**Production Date:** 16 Sep 2011

**Expiration Date:** 07 Sep 2013

**Primary Application:** Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

4  
4  
4

**Sample:** Carbamate Test Mixture, PN: 1700-0063.

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:**

*Jim Murphy*  
VP OPR  
Title

**Date:**

21 Sep 2011

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. A, 04/08/2011

P:\Documentation\Chemical Documentation\CB130